

Rev 09

Growing Global

## **EU Declaration of Conformity**

Manufacturer	:	Hartalega NGC Sdn. Bhd.
Manufacturer's Address	:	No.1, Persiaran Tanjung, Kawasan Perindustrian Tanjung, 43900 Sepang, Selangor, Malaysia.
EU Representative	:	Medical Device Safety Service (MDSS) Schiffgraben 41, 30175 Hannover, Germany.
Product Description (MDR)	:	Nitrile Powder Free Examination Gloves
Intended Purpose (MDR)		Nitrile Powder Free Examination Gloves are intended to be used to contribute to prevent cross contaminations in the framework of medical examinations and diagnostic/ therapeutic procedures conducted under non-sterile conditions.
Device Classification	:	Class I, according to Annex VIII of Regulation (EU) 2017/745
Rule (s)	:	1 and 5
Conformity Assessment Procedure	:	Annex II and Annex III
Basic UDI-DI	:	955100777HNGCTFMD005AQ8
Authorised Representative SRN	:	DE-AR-000005430
Manufacturer SRN	:	MY-MF-000010459
Reference to Trade Name (MDR)	ċ	Attachment I
Standard Reference (MDR)		Attachment II
Product Description (PPER)		Nitrile Powder Free Examination Gloves (≥ 2.2 mil)
		Available in standard minimum 240mm length or a longer cuff variant of 280mm
Device Classification (PPER)	÷	Category III (Type B)
EU Type-Examination Certificate Number (PPER)		2777/11578-02/E00-00
Reference to Trade Name (PPER)	:	Attachment III
Standard Reference (PPER)	:	EN 420:2003+A1:2009
		EN ISO 374-1:2016+A1:2018
		EN ISO 374-5:2016

Hartalega Holdings Berhad (741883-X)

C-G-9, Jalan Dataran SD1, Dataran SD PJU 9 Bandar Sri Damansara 52200 Kuala Lumpur, Malaysia Tel: +603 - 6277 1733 Fax: +603 - 6280 2533 www.hartalega.com.my Hartalega NGC Sdn Bhd (984586-P)

No. 1, Persiaran Tanjung Kawasan Perindustrian Tanjung 43900 Sepang, Selangor, Malaysia Tel: +603 - 8707 3000



We, Hartalega NGC Sdn. Bhd. herewith declared that above mentioned device:

- is in conformity with the Regulation (EU) 2017/745 of The European Parliament and of The Council of medical devices.
- is in conformity with the provisions of Regulation (EU) 2016/425 on personal protective equipment.
- is subject to the conformity assessment procedure Module C2 set out in Annex VII of Regulation (EU) 2016/425, under the surveillance of the notified body SATRA Technology Europe Limited, Bracetown Business Park, Clonee D15YN2P, Republic of Ireland (Notified Body number 2777).

This EU declaration of conformity is issued under the sole responsibility of the manufacturer, Hartalega NGC Sdn. Bhd.

Place and Date of Issue

: Hartalega NGC Sdn. Bhd./ 02<sup>nd</sup> March 2022

Signed for and on Behalf of Hartalega NGC : Sdn. Bhd.

Name : NURUL AISYAH KONG Position : DEPUTY GENERAL MANAGER – QUALITY ASSURANCE

Hartalega Holdings Berhad (741883-X)

C-G-9, Jalan Dataran SD1, Dataran SD PJU 9 Bandar Sri Damansara 52200 Kuala Lumpur, Malaysia Tel: +603 - 6277 1733 Fax: +603 - 6280 2533 www.hartalega.com.my Hartalega NGC Sdn Bhd (984586-P)

No. 1, Persiaran Tanjung Kawasan Perindustrian Tanjung 43900 Sepang, Selangor, Malaysia Tel: +603 - 8707 3000

Growing Global

Rev 09

## ATTACHMENT I

Product or Trade Name	<b>Reference Number</b>
	XS: 444XS
	S: 444S
PEPPLER NITRIL BLACK 444	M: 444M
	L: 444L
	XL: 444XL

No. of Concession, Name

## ATTACHMENT II

Standard	Title
ISO 9001:2015	Quality Management Systems – Requirements
EN ISO 13485:2016	Medical Devices – Quality Management Systems – Requirements for Regulatory Purposes
EN 455-1:2000	Medical Gloves for Single Use Part 1: Requirements and Testing for Freedom from Holes
EN 455-1:2020	Medical Gloves for Single Use Part 1: Requirements and Testing for Freedom from Holes
EN 455-2:2015	Medical Gloves for Single Use Part 2: Requirements and Testing for Physical Properties
EN 455-3:2015	Medical Gloves for Single Use Part 3: Requirements and Testing for Biological Evaluation
EN 455-4:2009	Medical Gloves for Single Use Part 4: Requirements and Testing for Shelf Life Determination
BS EN 1041:2008+A1:2013	Information Supplied by the Manufacturer of Medical Devices
BS EN ISO 14971:2019	Medical Devices – Application of Risk Management to Medical Devices
ISO 15223-1:2016	Medical Devices – Symbols to be Used with Medical Device Labels, Labelling and Information to be Supplied Part 1: General Requirements
ISO 10993-1:2018	Biological Evaluation of Medical Devices Part 1: Evaluation and Testing within a Risk Management Process
ISO 10993-5:2009	Biological Evaluation of Medical Devices Part 5: Tests for In Vitro Cytotoxicity
ISO 10993-10:2010	Biological Evaluation of Medical Devices Part 10: Tests for Irritation and Skin Sensitization
ISO 10993-11:2017	Biological Evaluation of Medical Devices Part 11: Tests for Systemic Toxicity
ISO 10993-18:2005	Biological Evaluation of Medical Devices Part 18: Chemical Characterization of Materials
ISO 2859- 1:1999/Amd1:2011	Sampling Procedures for Inspection by Attributes Part 1: Sampling Schemes Indexed by Acceptance Quality Limit (AQL) for Lot-By-Lot Inspection

## ATTACHMENT III

Product or Trade Name	<b>Reference Number</b>
PEPPLER NITRIL BLACK 444	XS: 444XS S: 444S M: 444M L: 444L XL: 444XL